

Central Laboratories Department  
Food Laboratory Section

إدارة المختبرات المركزية  
قسم مختبر الأغذية

## Test Report

Ref.No	F2401971	Received Date	20/03/2024	Date of Issue	27/03/2024
Client & Sample Details					
Name :	WOODLEM PARK SCHOOL				
Source :	WOODLEM PARK SCHOOL				
Sample :	Drinking Water	Certificate No. :			
Description :	K.G Section	Sampling by :			
Brand :	NO Brand	Packaging Type:	Plastic Bottle		
Net wt/volume per unit :	/	Origin :	UAE		
Reason of Analysis :	School Survey	Prod.Date :			
Remarks :	Qty:-02Pcs(1Plastic+1Glass) /Sample Condition:-Room Temp(22.8c°)/NO Barcode & Batch Number / Time:-10:30AM				

Parameter	Method	Result	Unit	Ref. Limit	Remarks
<b>PHYSICAL RESULTS</b>					
General appearance	Analysis Start Date	20/03/2024	Analysis End Date	20/03/2024	
Label and Packaging	U.A.E std 9 , 150	/			
Organoleptic Evaluation		Unobjectionable			

### Satisfactory

Parameter	Method	Result	Unit	Ref. Limit	Remarks
<b>MICROBIOLOGICAL</b>					
E.coli	APHA 9223 B/IDEXX manual	Not Detected	/100 ml	Not Detected (i)	
Pseudomonas aeruginosa	APHA 9213 G / IDEXX manual	Not Detected	/250 ml	Not Detected (i)	
Total Coliforms	APHA 9223 B/IDEXX manual	Not Detected	/100 ml	Not Detected (i)	

### Satisfactory

Parameter	Method	Result	Unit	Ref. Limit	Remarks
<b>CHEMICAL RESULTS</b>					
Free Chlorine		< 0.2	ppm		
Conductivity	APHA 2510 B/EC meter	221	us/cm		
TDS	APHA 2510 B/EC meter	133	mg/L	(100-1000)	
pH	APHA 4500 H+ B/pH meter	7.1	pH Unit	6.5 - 8.5	
Bromate	IC	< 5	ppb	=<10	

### Satisfactory

Final Result	<b>Investigated Parameters are Compliant with Regulations/Standards</b>				
General Comments	Micro Results: UAE.S1016:2017. chemistry result : UAE.S GSO 1025 : 2014				

حكومة عجمان | إدارة البلدية والتخطيط  
Government of Ajman | Municipality & Planning Department  
إدارة المختبرات المركزية  
Central Laboratories Dept.  
النتائج متوافقة مع اللوائح / المعايير  
Investigated Parameters are  
Compliant with Regulations/Standards



## Test Report

Ref.No	F2401970	Received Date	20/03/2024	Date of Issue	27/03/2024
--------	----------	---------------	------------	---------------	------------

### Client & Sample Details

Name :	WOODLEM PARK SCHOOL	Certificate No. :	
Source :	WOODLEM PARK SCHOOL	Sampling by :	
Sample :	Drinking Water	Packaging Type:	Plastic Bottle
Description :	Primary Section	Origin :	UAE
Brand :	NO Brand	Prod.Date :	
Net wt/volume per unit :	/	Exp.Date :	
Reason of Analysis :	School Survey	Exp.Date :	
Remarks :	Qty:-02Pcs(1Plastic+1Glass) /Sample Condition:-Room Temp(22.8c°)/NO Barcode & Batch Number / Time:-10:30AM		

### Test Result

Parameter	Method	Result	Unit	Ref. Limit	Remarks
<b>PHYSICAL RESULTS</b>					
Analysis Start Date		20/03/2024	Analysis End Date		20/03/2024
General appearance		Satisfactory			
Label and Packaging	U.A.E std 9 , 150	/			
Organoleptic Evaluation		Unobjectionable			

### Satisfactory

<b>MICROBIOLOGICAL</b>					
Analysis Start Date		20/03/2024	Analysis End Date		21/03/2024
E.coli	APHA 9223 B/IDEXX manual	Not Detected	/100 ml	Not Detected (i)	
Pseudomonas aeruginosa	APHA 9213 G / IDEXX manual	Not Detected	/250 ml	Not Detected (i)	
Total Coliforms	APHA 9223 B/IDEXX manual	Not Detected	/100 ml	Not Detected (i)	

### Satisfactory

<b>CHEMICAL RESULTS</b>					
Analysis Start Date		20/03/2024	Analysis End Date		21/03/2024
Free Chlorine		< 0.2	ppm		
Conductivity	APHA 2510 B/EC meter	218	us/cm		
TDS	APHA 2510 B/EC meter	131	mg/L	(100-1000)	
pH	APHA 4500 H+ B/pH meter	7.2	pH Unit	6.5 - 8.5	
Bromate	IC	< 5	ppb	=<10	

### Satisfactory

Final Result	Investigated Parameters are Compliant with Regulations/Standards
General Comments	Micro Results: UAE.S1016:2017. chemistry result : UAE.S GSO 1025 : 2014



## Test Report

Ref.No	F2401970	Received Date	20/03/2024	Date of Issue	27/03/2024
--------	----------	---------------	------------	---------------	------------

### Client & Sample Details

Name :	WOODLEM PARK SCHOOL				
Source :	WOODLEM PARK SCHOOL	Certificate No. :			
Sample :	Drinking Water	Sampling by :			
Description :	Primary Section	Packaging Type:	Plastic Bottle		
Brand :	NO Brand	Origin :	UAE		
Net wt/volume per unit :	/	Prod.Date :			
Reason of Analysis :	School Survey	Exp.Date :			
Remarks :	Qty:-02Pcs(1Plastic+1Glass) /Sample Condition:-Room Temp(22.8c°)/NO Barcode & Batch Number / Time:-10:30AM				

### Test Result

Parameter	Method	Result	Unit	Ref. Limit	Remarks
<b>PHYSICAL RESULTS</b>					
Analysis Start Date		20/03/2024	Analysis End Date		20/03/2024
General appearance		Satisfactory			
Label and Packaging	U.A.E std 9 , 150	/			
Organoleptic Evaluation		Unobjectionable			

### Satisfactory

<b>MICROBIOLOGICAL</b>					
Analysis Start Date		20/03/2024	Analysis End Date		21/03/2024
E.coli	APHA 9223 B/IDEXX manual	Not Detected	/100 ml	Not Detected (i)	
Pseudomonas aeruginosa	APHA 9213 G / IDEXX manual	Not Detected	/250 ml	Not Detected (i)	
Total Coliforms	APHA 9223 B/IDEXX manual	Not Detected	/100 ml	Not Detected (i)	

### Satisfactory

<b>CHEMICAL RESULTS</b>					
Analysis Start Date		20/03/2024	Analysis End Date		21/03/2024
Free Chlorine		< 0.2	ppm		
Conductivity	APHA 2510 B/EC meter	218	us/cm		
TDS	APHA 2510 B/EC meter	131	mg/L	(100-1000)	
pH	APHA 4500 H+ B/pH meter	7.2	pH Unit	6.5 - 8.5	
Bromate	IC	< 5	ppb	=<10	

### Satisfactory

Final Result	Investigated Parameters are Compliant with Regulations/Standards				
General Comments	Micro Results: UAE.S1016:2017. chemistry result : UAE.S GSO 1025 : 2014				

